

## Werkzeugzeugnis Test report

Das Werkprüfzeugnis wurde erstellt auf Grundlage DIN EN 10204 Punkt 2.2.  
This report has been made in accordance with DIN EN 10204 item 2.2.

Hiermit bescheinigen wir, dass dieses Produkt in Übereinstimmung mit dem QUM-Handbuch der Feingerätebau K. Fischer GmbH nach DIN EN ISO 9001 und DIN EN ISO 14001 gefertigt wurde. Die Ausführung und Anzeigegenauigkeit der Geräte / Systeme wurde im Rahmen der Qualitätssicherungsmaßnahmen überwacht.  
Die Qualitätsprüfung ergab keine Beanstandungen.

This is to certify, that this product has been tested according to the integrated management system of the Feingerätebau K. Fischer GmbH manual in accordance with DIN EN ISO 9001 and DIN EN ISO 14001. Execution of instruments / systems as well as testing of accuracy was carried out following quality assurance procedures.  
Quality inspection was successfully passed.

Erzeugnis Product	Präzisions-Aneroidbarometer
Geräte Typ Model type	103PM
Serien Nr. Serial Nr.	2758
Kunde Customer	

	T [K]	r.F. [%] Bei ____ °C	P [hPa] Bei <u>20</u> °C
Genauigkeit Accuracy	-----	-----	±0,7 (890...1050)

Prüfer  
Checked by

12.10.2009  
Datum/Date

  
Unterschrift/Signature

Qualitätssicherung  
Quality control

12.10.2009  
Datum/Date

  
Unterschrift/Signature